

SAIP2013



Contribution ID : 300

Depletion voltage measurements of the iThemba LABS segmented clover detector.

Tuesday 09 Jul 2013 at 16:20 (00h20')

Abstract :

J.L. Easton^{1,2}, O. Shirinda¹, E.A. Lawrie¹, T. D. Bucher¹, S.P. Noncolela^{1,2}, N. Orce^{2,1} iThemba LABS, PO Box 722, 7129 Somerset West, South Africa² University of the Western Cape, Private Bag X17, 7535 Bellville, South Africa The measured depletion voltages of the four crystals of the iThemba LAB segmented clover detector was compared to the manufacturer specifications. We measured the depletion voltage by measuring the peak centroids and peak areas as a function of applied voltage. We had used three sources namely ¹³⁷Cs, ⁶⁰Co and ²⁴¹Am. A reduced chi squared analysis was then used to infer depletion voltages. It was found that the depletion voltage was higher than the manufacturer specifications for all four crystals. The depletion voltages depend on the amount of the impurities in each crystal, which are very important for simulating correctly the gamma-ray interaction points inside the detector.

Award :

yes

Level :

PhD

Supervisor :

E.A. Lawrie¹, elena@tlabs.ac.za

Paper :

no

Primary authors : Mr. EASTON, Jayson (iThemba LABS and University of the Western Cape)

Co-authors :

Presenter : Mr. EASTON, Jayson (iThemba LABS and University of the Western Cape)

Session classification : NPRP

Track classification : Track B - Nuclear, Particle and Radiation Physics

Type : Oral Presentation