



Contribution ID: 293

Type: Oral Presentation

## In-situ neutron powder diffraction temperature capabilities at SAFARI-1

Wednesday, 1 July 2015 11:50 (20 minutes)

**Abstract content (Max 300 words)** [http://events.saip.org.za/getFile.py/?target=\\_blank](http://events.saip.org.za/getFile.py/?target=_blank) **Formatting & Special chars**

The recently commissioned neutron powder diffractometer at the SAFARI-1 reactor of Necsa, aptly named Powder Instrument for Transition in Structure Investigations (PITSI), is equipped with both heating and cooling sample environments to allow in-situ temperature dependent studies in the temperature range 3 K to 1800 K. SAFARI-1 is a 20 MW light-water moderated tank-in-pool materials testing reactor that has an in-core flux of  $4 \times 10^{14}$  neutrons  $\text{cm}^{-2}\text{s}^{-1}$ . The monochromatic flux of the PITSI instrument at the sample position is  $\approx 106$  neutrons  $\text{cm}^{-2}\text{s}^{-1}$ . The interchangeable heating and cooling sample environments were both procured from AS Scientific Products Ltd., UK and are specifically configured for use on neutron scattering instruments. The closed-cycle cryostat is equipped with a SHI Cryogenics RDK-408D2 cold head rendering 1W cooling power at 3.5 K, which in conjunction with a LakeShore 340 temperature controller, facilitates achieving accurate temperatures covering the range  $3 \text{ K} < T < 350 \text{ K}$ . The vacuum furnace employs radiant heat generated and focussed with a 3600 W Niobium element and heat shrouds in conjunction with a West 4100+ temperature controller covers temperatures from 300 K to 1800 K. Both these sample environments are under full computer control. Temperature set points can be maintained to accuracies better than  $\pm 1 \text{ K}$  and logged throughout the data acquisition. With envisaged modifications, the furnace system could be extended to incorporate in-situ reaction chambers.

PITSI in conjunction with its temperature environments, amongst others enable in-situ studies of chemical and / or magnetic phase transitions, determination of thermal expansion coefficients, quantification of phases, etc. A number of investigations have been completed and will be reported on to demonstrate the capabilities of this instrument that is available to the research community under the R&D User Program.

**Apply to be considered for a student award (Yes / No)?**

no

**Level for award (Hons, MSc, PhD, N/A)?**

no

**Would you like to submit a short paper for the Conference Proceedings (Yes / No)?**

yes

**Please indicate whether<br>this abstract may be<br>published online<br>(Yes / No)**

yes

**Primary author:** Dr HAYES, Michael (Necsa)

**Co-authors:** Dr VENTER, Andrew (Necsa); Mr RAATHS, Christo (Necsa); Mr MARAIS, Deon (Necsa); Ms NELWAMONDO, Maulusi (Necsa); Mr VAN HEERDEN, Rudolph (Necsa); Ms SENTSHO, Zeldah (Necsa)

**Presenter:** Dr HAYES, Michael (Necsa)

**Session Classification:** Applied

**Track Classification:** Track F - Applied Physics