



Contribution ID: 380

Type: Oral Presentation

## Measurement of the cosmogenic isotope $^{10}\text{Be}$ at iThemba LABS and applications of cosmogenic isotopes in South Africa

Thursday, 28 June 2018 10:40 (20 minutes)

$^{10}\text{Be}$  is an important isotope for accelerator mass spectrometry system of iThemba LABS in Johannesburg, because of local demand for cosmogenic radionuclide dating methods in the local earth science and paleosciences community. This demand also meets a very suitable basic AMS setup with its ion-source based on a version of the Cesium Sputter source of the Lawrence Livermore National Laboratories, which delivers significantly higher output of  $\text{BeO}^-$  than commercially available ion sources. The measurement of  $^{10}\text{Be}$  by accelerator mass spectrometry requires the effective suppression of  $^{10}\text{B}$  before the detector, which is either done by using a gas absorber cell for full stopping of  $^{10}\text{B}$ , or by differential energy loss of  $^{10}\text{B}$  versus  $^{10}\text{Be}$  in an absorber foil (post-stripping) followed by a high-resolution magnet or electrostatic analyser. The latter method has proven effective in that the  $2+$  charge state can be utilized, which has in excess of 40% charge state transmission at 4MV terminal voltage, but there are charge state losses from the absorber foil. Using gas absorber cell employing havar windows usually requires higher energies, requiring the  $3+$  charge state and higher terminal voltage, resulting in additional interference from the  $^{10}\text{B}(p,)^7\text{Li}$  reaction that is not fully separated in the detector. Recently it has been shown that low stress silicone nitride membranes can be used as absorber for full stopping with particle energies of 6MeV for  $^{10}\text{Be}$ . This allows for the use of the  $2+$  charge state, avoids the charge state losses of the post-stripping method, and – provided the chosen terminal voltage is low enough – it avoids incurring massive interference from the nuclear reaction. We implemented this method in lieu of the traditional gas absorber cell, thus cashing in on the efficiency gain from using the  $2+$  charge state. Together with the efficiency benefits of our ion-source we have a high detection efficiency AMS system for  $^{10}\text{Be}$ . We applied the foilstack methods successfully in test runs and recently in first projects of our cosmogenic isotope programme.

**Please confirm that you have carefully read the abstract submission instructions under the menu item "Call for Abstracts" (Yes / No)**

Yes

**Consideration for student awards** Choose one option from those below.  
N/A  
Hons  
MSc  
PhD

N/A

**Supervisor details** If not a student, type N/A.  
Student abstract submission requires supervisor permission: please give their name, institution and email address.

N/A

**Primary author:** Dr WINKLER, Stephan (NRF/iThemba LABS)

**Co-authors:** Prof. KRAMERS, Jan (University of Johannesburg); Mrs KHOSA, Rivoningo (University of Johannesburg); Dr MULLINS, Simon (NRF/iThemba LABS); Mr MAKHUBELA, Tebogo (University of Johannesburg); Dr MBELE, Vela (iThemba LABS)

**Presenter:** Dr WINKLER, Stephan (NRF/iThemba LABS)

**Session Classification:** Nuclear, Particle and Radiation Physics

**Track Classification:** Track B - Nuclear, Particle and Radiation Physics