



UNIVERSITEIT VAN PRETORIA
UNIVERSITY OF PRETORIA
YUNIBESITHI YA PRETORIA

Contribution ID: 67

Type: **Poster Presentation**

Photon dose distributions between 3D-water phantom and Profiler2 scanning system

Tuesday, 10 July 2012 17:30 (2 hours)

Abstract content
 (Max 300 words)

In this study we aim to determine whether a Profiler2 scanning system can be used as a substitute for 3D-water phantom used during commissioning of linear accelerator. All the measurements were performed with 6 and 15 MV photon beams generated by Elekta Synergy linear accelerator. Percentage depth-dose and beam profiles were measured for the 6×6, 10×10, 14×14, 20×20 and 25×25 cm² field sizes defined at 0.5, 1.0, 2.0 and 5.0 cm in 3D-water phantom and Pespex. Dose distributions compared well within recommended limits with the maximum difference in symmetry of 1.6%, flatness of 2.5 mm, penumbra of -12.11 mm, which was outside the recommended limit of 12 mm, and PDD10 of 1.69%. The results of validation of both 3D-water phantom and Profiler2 scanning system with modeling of data using XiO TPS system, showed an agreement within 1%. The results showed the suitability of the profiler2 scanning system for commission of linear accelerator.

Apply to be
 consider for a student
 award (Yes / No)?

No

Would you like to
 submit a short paper
 for the Conference
 Proceedings (Yes / No)?

No

Primary author: Dr SITHOLE, Enoch (member)

Presenter: Dr SITHOLE, Enoch (member)

Session Classification: Poster Session

Track Classification: Track B - Nuclear, Particle and Radiation Physics