



Contribution ID: 79

Type: Poster Presentations

## A Widely Tunable 10- $\mu\text{m}$ QCL Locked to a Metrological Mid-IR Reference for Precision Molecular Spectroscopy

Tuesday, 29 January 2019 16:35 (5 minutes)

The quantum cascade lasers (QCL) are popular sources for spectroscopy in the field of mid-infrared because of the wide range of wavelengths they can cover ( $3\ \mu\text{m} < \lambda < 24\ \mu\text{m}$ ). Several examples of spectroscopic measurements with spectrometers based on QCL, have been demonstrated [2].

We are currently developing a laser spectrometer based on a QCL which emits around  $10\ \mu\text{m}$ . The selection of this wavelength for the QCL source is to compare it to our existing ultra-stable  $\text{CO}_2$  lasers. We characterized a free-running continuous wave near-room-temperature distributed feedback  $10.3\ \mu\text{m}$  QCL. This gave a remarkable result on the frequency noise which is an order of magnitude smaller compared to what was published on the characterization of these types of lasers sources. A full width at half maximum (FWHM) equal to  $60\ \text{kHz}$  of the beat signal between the free-running QCL and a  $1\text{-kHz}$  narrow  $\text{CO}_2$  laser was observed after  $1\ \text{ms}$  of integration time.

Narrowing of the QCL line width has been made by taking a phase-locked QCL on the  $\text{CO}_2$  laser which is itself stabilized on a saturated absorption transition of the  $\text{OsO}_4$  molecule. The beat spectrum between phase-locked QCL and  $\text{CO}_2$  laser recorded with a radio-frequency (RF) spectrum analyzer allowed us to estimate that more of 99% of the beat signal RF power is concentrated in the carrier. This allows to conclude that the QCL was copied almost exactly the spectral characteristics of our ultra-stable  $\text{CO}_2$  laser ( $10\text{-Hz}$  line width, accuracy of a few tens of hertz). These results in a record QCL line width of the order of  $10\ \text{Hz}$ , 3 to 4 orders of magnitude lower than a free-running QCL, and a relative stability at  $1\ \text{s}$  of about  $1\ \text{Hz}$ .

The phase-locked QCL was then used to measure the spectra of ammonia ( $\text{NH}_3$ ) and methyltrioxorhenium (MTO) to demonstrate its potential for two main projects of our group: the determination of the Boltzmann constant,  $k_B$ , by Doppler spectroscopy of ammonia [3] and the first observation of parity violation by Ramsey interferometry of a beam of chiral molecules [4].

### References

- [1] Kosterev et al., Applied Physics B, 90(2), 165-176 (2008)
- [2] Wysocki et al., Applied Physics B, 92(3), 305-311 (2008)
- [3] C. Lemarchand et al., Metrologia 50 623 (2013)
- [4] S. K. Tokunaga et al., Molecular Physics 111, Issue 14-15, 2363–2373 (2013)

**Primary author:** Dr SOW, Papa Lat Tabara (Université Alioune DIOP de Bambey)

**Co-authors:** Prof. GONCHAROV, Andrey (Institute of Laser Physics of SB RAS, Novosibirsk); Prof. AMY-KLEIN, Ane (Université Paris 13, France); Dr DARQUIÉ, Benoît (Université Paris 13, France); Dr ARGENCE, Bérengère (Université Paris 13, France); Prof. CHARBONNET, Christian (Université Paris 13, France); Prof. DAUSSY, Christophe (Université Paris 13, France); Mr LOPEZ, Olivier (Université Paris 13, France); Dr TOKUNAGA, Sean (Université Paris 13, France); Dr MEJRI, Sinda (Université Paris 13, France)

**Presenter:** Dr SOW, Papa Lat Tabara (Université Alioune DIOP de Bambey)

**Session Classification:** AFLS2 Poster Session

**Track Classification:** AFLS2 track