SAIP2015



Contribution ID: 330

Type: Poster Presentation

Orientation of the Ge crystals of the iThemba LABS segmented clover detector

Wednesday, 1 July 2015 16:10 (1h 50m)

Abstract content
 (Max 300 words)
Formatting &
Special chars

The iThemba LABS segmented detector include four Ge crystals in a clover configuration in a single cryostat [1]. The orientation of the Ge crystals inside the cryostat is measured with a collimated source, where the axis of the collimator is perpendicular to the front face of the detector cryostat.

In addition, the orientation of the crystal lattice is being investigated by measuring the T30 and T90 rise times. This orientation plays a crucial role in the correct simulation of the charge collection and generating realistic pulse shapes.

[1] F. A. Beck, et al., Proceedings of Workshop on Large g Ray Detector Arrays, Chalk River, AECL-10613. 1994, p. 359.

Apply to be
 considered for a student
 award (Yes / No)?

No

Level for award
 (Hons, MSc,
 PhD, N/A)?

N/A

Main supervisor (name and email)
and his / her institution

Elena Lawrie elena@tlabs.ac.za iThemba LABS

Would you like to
 submit a short paper
 for the Conference
 Proceedings (Yes / No)?

Yes

Please indicate whether
this abstract may be
published online
(Yes / No)

Yes

Primary author: Dr DINOKO, Tshepo (iThemba LABS)

Co-authors: Dr BUCHER, Daphney (iThemba LABS); Dr LAWRIE, Elena (iThemba LABS); Mr EASTON, Jayson (iThemba LABS); Mr ERASMUS, Nicolas (iThemba LABS); Dr SHIRINDA, Obed (iThemba LABS); Mr NONCOLELA, Sive (iThemba LABS)

Presenter: Dr DINOKO, Tshepo (iThemba LABS)

Session Classification: Poster2

Track Classification: Track B - Nuclear, Particle and Radiation Physics