



Contribution ID: 348

Type: Poster Presentation

A high speed OCT system developed at the CSIR National Laser Centre

Wednesday, 6 July 2016 16:10 (1h 50m)

Abstract content (Max 300 words)Formatting & Special chars

Light based techniques continue to gain momentum in different spheres of diagnostic and therapeutic applications as a result of their non-invasive, non-contact properties. One such technique is Optical Coherence tomography (OCT). Since it was first reported by Huang in 1991[1], OCT has made significant strides in different fields from dermatology and ophthalmology to polymer characterisation and bio-metrics[2-4]. In South Africa, the technique is still emerging although it is being used for eye examinations by ophthalmologists. The type of OCT system employed can be a simple, cost effective solution or a complex, highly specific and fast system depending on the application.

As part of a larger project, the CSIR National Laser Centre has designed and built a high speed OCT system that can image a large surface area (25 by 25 mm) to a depth of 11 mm (sample dependant). Resultant 3-D images (512 x 512 x 2048 pixels) are acquired in less than 3 seconds. The performance of the system compares adequately with many commercially available systems which usually image smaller areas [5-6].

The heart of the system is a 200 kHz swept laser source and two axis galvanometer based scanner. Signal acquisition is made possible through a high speed analogue-to-digital converter capable of speeds greater than 1GS/s. This paper will give an overview of the system and elaborate on the design of the data acquisition system and the initial results that have been obtained.

Apply to be considered for a student award (Yes / No)?

No

Level for award (Hons, MSc, PhD, N/A)?

N/A

Main supervisor (name and email) and his / her institution

N/A

Would you like to submit a short paper for the Conference Proceedings (Yes / No)?

No

**Please indicate whether
this abstract may be
published online
(Yes / No)**

Yes

Primary author: Mr SHARMA, Ameeth (CSIR National Laser Centre)

Co-authors: Ms SINGH, Ann (CSIR,National Laser Centre); Dr STRAUSS, Hencharl (CSIR (National Laser Centre)); Mr RAMOKOLO, Rocky (CSIR, National Laser Centre); Dr ROBERTS, Ted (CSIR National Laser Centre)

Presenter: Mr SHARMA, Ameeth (CSIR National Laser Centre)

Session Classification: Poster Session (2)

Track Classification: Track F - Applied Physics