
SAIP2016



Contribution ID: 428

Type: Poster Presentation

A comparison of neutron energy distributions unfolding codes used with a NE213 detector

Wednesday, 6 July 2016 16:10 (1h 50m)

Abstract content
 (Max 300 words)
 dry-Formatting &
 &classed chars

In fast neutron radiotherapy knowledge of how neutrons transfer energy into body tissue as they progress through the body is required. This knowledge is important to calculate the energy distributions of secondary charged particles, to characterize the radiation quality and absorbed dose both inside and near to the area under treatment. Experimental measurement or Monte Carlo transport codes can be used to determine neutron energy distributions. Unfolding can be used for processing the measured data and obtaining the neutron energy distribution. The neutron energy distributions have to be unfolded from the measured pulse-height distributions, which is an ill-conditioned problem. Therefore, use of independent unfolding methods allows for comparison and interpretation of the data.

This paper reports on the overall procedure of measuring and unfolding the fast neutron energy distributions with NE213 liquid scintillation detector using the unfolding codes Gravel based on the GRAVEL iterative method and MAXED based on the Maximum Entropy method. The results obtained from pulse height distributions using the different unfolding codes will be presented and discussed.

Apply to be
br> considered for a student
br> award (Yes / No)?

No

Level for award

- (Hons, MSc,

- PhD, N/A)?

No

Would you like to
 submit a short paper
 for the Conference
 Proceedings (Yes / No)?

yes

Please indicate whether

br>this abstract may be

br>published online

(Yes / No)

yes

Primary author: Dr HERBERT, Mark (University of the Western Cape) **Co-author:** Mr SHILUVANE, Thulani (University of the Western Cape)

Presenter: Dr HERBERT, Mark (University of the Western Cape)

Session Classification: Poster Session (2)

Track Classification: Track F - Applied Physics