



Contribution ID: 230

Type: Oral Presentation

Measurements of natural radioactivity in sands using an array of lanthanum -bromide scintillator detectors

Wednesday, 5 July 2017 15:00 (20 minutes)

LaBr₃:Ce detectors have been shown to be 1.2–1.65 times more efficient than NaI:Tl detectors above 350 keV, for 3.8 cm×3.8 cm (1.5 in.×1.5 in.) detectors and have an energy resolution of 2.5–3% at the 662 keV gamma-line of ¹³⁷Cs, compared to 6–7% for NaI:Tl detectors[1]. The detector crystal has other advantages such as a high scintillation light output with a fast decay time[2]. An array of 8 2in x 2in LaBr₃(Ce) scintillators with an XIA PIXIE-16 Digital Signal Processing system data acquisition system will be used to measure sands and KCl sample placed in the centre (24cm from each detector) of the array (with all detectors lying in the horizontal plane) for 12 hours. The gamma-gamma coincidence method has the advantage of virtually eliminating all background peaks that do not exist in coincidence with other peaks, significantly improving detection limits of useful radionuclides[3][4]. By employing a gamma-gamma coincidence condition, the background from the radioisotopes in the LaBr₃:Ce scintillator is eliminated, providing a means for improving detection limits[5]. The absolute gamma-ray energy detection efficiency of each detector will be determined and compared. Data from each detector will be analyzed. The activity concentration of ²³⁸U, ²³²Th and ⁴⁰K in the sands will be determined and compared to certified values. Time-stamped data will be collected and then coincidence conditions between detectors set offline. In this way this work can make a comparison between traditional single measurements and coincidence method.

Summary

Reference

- [1] K. Ciupek, S. Jednoróg, M. Fujak, and K. Szewczak, "Evaluation of efficiency for in situ gamma spectrometer based upon cerium-doped lanthanum bromide detector dedicated for environmental radiation monitoring," J. Radioanal. Nucl. Chem., vol. 299, no. 3, pp. 1345–1350, 2014.
- [2] A. Favalli, H. C. Mehner, and F. Simonelli, "Wide energy range efficiency calibration for a lanthanum bromide scintillation detector," Radiat. Meas., vol. 43, no. 2–6, pp. 506–509, 2008.
- [3] M. Yoho and S. Landsberger, "Determination of Selenium in coal fly ash via γ - γ coincidence neutron activation analysis," J. Radioanal. Nucl. Chem., vol. 307, no. 1, pp. 733–737, Jan. 2016.
- [4] S. Horne and S. Landsberger, "Selenium and mercury determination in biological samples using gamma-gamma coincidence and Compton suppression," J. Radioanal. Nucl. Chem., vol. 291, no. 1, pp. 49–53, Jan. 2012.
- [5] A. Drescher et al., "Gamma-gamma coincidence performance of LaBr₃:Ce scintillation detectors vs HPGe detectors in high count-rate scenarios," Appl. Radiat. Isot., vol. 122, no. January, pp. 116–120, 2017.

Apply to be considered for a student award (Yes / No)?

yes

Level for award (Hons, MSc, PhD, N/A)?

PhD

**Main supervisor (name and email)
and his / her institution**

Prof. R. T. Newman, rtnewman@sun.ac.za, Stellenbosch University

**Would you like to
 submit a short paper
 for the Conference
 Proceedings (Yes / No)?**

yes

Primary author: Ms BASHIR, M (Stellenbosch University and Ibrahim Badamasi Babangida University)

Co-authors: Dr JONES, P (iThemba Laboratory for Accelerator Based Sciences (iThemba LABS)); Prof. NEWMAN, R. T (Stellenbosch University)

Presenter: Ms BASHIR, M (Stellenbosch University and Ibrahim Badamasi Babangida University)

Session Classification: Nuclear, Particle and Radiation Physics 1

Track Classification: Track B - Nuclear, Particle and Radiation Physics